



## Instruction for Use – Cytology Brushes SU

### Attention:

**This medical device may only be purchased by specialists, doc-tors and medical personnel and may only be used in accordance with these instructions for use and the defined area of application.**

### Description / Versions

ENDO-FLEX GmbH offers the following versions:

- Double lumen cytology brushes Ø 2.7 mm, Length 180 cm, for use with guide wire
- Single lumen cytology brushes Ø 1.8 and 2.3 mm, Length 120 to 230 cm, for use without guide wire

### Products

Single Lumen:	Double Lumen:
E4218-A	E4222DL-C
E4218-C	
E4219-G	
E4222-G	

### Important Note

Please ensure you read this user manual carefully before each use and keep it somewhere that is easily accessible for users and/or appropriate specialist staff.

**Ensure you carefully read through the warnings. Patients, users or third parties may suffer serious injury as a result of the improper use of the products.**

### Content and Packaging

- 1 outer box
- 5 Cytology Brushes SU (Single Use) individually sterile packed
- 1 Instructions for use

### Patient Population

The patient population or target patient group is derived from the indication determined by the responsible doctor who diagnostically or therapeutically treats the patient as part of an endoscopic procedure (the leading procedure per se) according to the intended use of the medical device.

There are no known restrictions to the patient population or the target patient group.

#### • Use of the product on minors:

The user can use the product on minors in the case that the physiological and anatomical conditions of the patient permit the use of the product.

#### • Use of the product on women who are pregnant or breastfeeding:

The indication for the use of the product on women who are pregnant or breastfeeding must be narrowed down by the user based on the respective individual physiological and anatomical conditions of the patient.

### 1. Scope of Application

The products listed above must only be used by physicians with appropriate specialist training in gastroenterology. The products are intended solely for the medical sector as illustrated below and must therefore be used within an operating environment suitable for this purpose. It is essential that both the user as well as the appropriate specialist staff familiarize themselves with the instruments before the user makes use of them.

### 2. Application Period

These instruments are intended to be used uninterrupted for a period of less than 60 minutes under normal conditions. (MDD 93/42 EEC)

### 3. Intended Use

Cytology Brushes are used for taking smears from the human body during endoscopic examinations, and to recover the collected cells and particles for further cytological examinations.

#### Indications

- histological cell probe collecting

#### Contraindications

- Not fasting patients
- Fragility of the intestinal wall: e.g. highly florid inflammation of the colon (e.g. ulcerative colitis, diverticulitis, ulcerative colitis, toxic megacolon)
- Peritonitis, acute abdomen, e.g. intestinal perforation, ileus
- Sepsis
- Co-morbidity, e.g. severe cardiopulmonary diseases and decompensation
- Uncontrollable haemorrhagic diatheses
- Pregnancy
- Recently created gastrointestinal anastomosis.

### 4. Complications / Side Effects / Cross-reactions

Sedation during endoscopic examination increases the risk of hypoxemia, hypercapnia, hypotension, arrhythmias and aspiration due to reduced protective reflexes. Hypoxemia also occurs without sedation during endoscopic examinations due to the feed of the endoscope.

In case of lesions of the pancreas, bleeding s were recorded occasionally. Due to the relative stiffness of the brushes for cell sampling, a bleeding lesion of the mucosa may not be excluded.

In case of bronchoscopy, brushing is usually related with minimal bronchial hemorrhage. However, in most cases, additional measures to control bleeding are not needed. Other possible complications are perforation and infection.

### 5. Warnings / Precautions

These instructions must be followed, as well as instructions from compatible components and hospital regulations for infection prevention, safe use, cleaning and sterilization. Failure to do so may result in serious injury to the patient and/or the user.

The following applies to the product:

- Sterile only if packaging is undamaged or unopened!
- For single use only! Do not reuse, reprocess or sterilize several times. Reuse, reprocessing or repeated sterilization of the instrument may affect its structural integrity and cause malfunction, resulting in contamination, infection and serious injury.
- If the instrument accidentally becomes dirty before treatment, it must be disposed of immediately! No cleaning agents may be applied.
- Do not use after the expiration date!
- All components should be carefully checked for compatibility and integrity before use. Do not use defective instruments! If defects occur, dispose of the instrument and replace it with a new one.
- Never use the product outside the recommended technical specifications (intended use).
- Never tamper with the structural conditions of the instrument, avoid kinks and other damage, and immediately discontinue use in the event of a malfunction!
- Wear protective clothing (gloves, goggles, gown, etc.) is absolutely necessary!
- Comparison of the technical data of the product with those of the endoscope used. The working channel diameter must be at least 0.2 mm larger than the outer diameter of the instrument.
- Never force instruments into the working channel!

### 6. Compatibility

Endoscopes / Guide Wires .025" to .035"

### 7. Function Test

The medical devices must be checked with regard to the following aspects prior to use:

- Expiration date
- Undamaged packaging
- Damage on the product (cracks on tube, bending, deformation)

Inspect products for immaculate surfaces, correct assembly and functionality.

**Products that have failed the function test must not be used as their sterility and product safety cannot be ensured. Dispose of these products accordingly, or please return them to the manufacturer.**

### 8. Preparation / Application

- Check the instrument for reliable function and irregularities before the procedure.
- Test the smooth movement of the instrument by carefully moving the finger slide forwards and backwards.
- **If you notice any irregularities, replace the instrument with a new one!**
- The brush is inserted into the working channel with the brush retracted into the tube. The corresponding working channel and tube diameters must be observed.
- 2-lumen cytology brushes allow insertion using a guide wire.
- Place the tube in front of the tissue to be sampled.
- Move the brush completely out of the tube into the fabric by carefully pushing the finger slide forward.
- Carefully pull back the finger slide to retract the brush, which is covered with tissue samples, completely back into the tube.
- Remove the cytology brush from the endoscope.

### 9. Sterility

#### Delivery condition

Disposable medical devices are supplied in an ETO gas-sterilized manner. Renewed preparation and sterilization following use is no longer possible and is prohibited! The product may only be used once.

Resterilisation should not be performed following expiry of the shelf life, i.e. the product must be disposed of according to clinical provisions.

### 10. Shelf Life of Products

The shelf life of the product is typically 3 years after the date of manufacture under normal conditions.

### 11. Liability and Warranty

As the manufacturer here, ENDO-FLEX shall not be liable for any damage or consequential damage arising from improper use or handling. This shall in particular apply to any use not in line with the defined purpose or failure to observe the preparation



## **Instruction for Use – Cytology Brushes SU**

and sterilisation instructions and warnings. This shall also apply to repairs or modifications to the product made by individuals who are not authorised to do so by the manufacturer.

### **12. Service**

Do not carry out any modifications to the product. If you have any concerns, complaints or comments regarding our products, please get in touch with us.

### **13. Transport and Storage Conditions**

- Products may only be transported and stored in the packaging provided for this purpose.
- Products must be stored dry and protected from sunlight at room temperature.
- Do not place any objects on the storage packaging and the sterile barrier system!
- Ensure the instruments are not kept near to chemicals, disinfectants or radioactive radiation!

### **14. Disposal**

After use, this product may present a biological hazard.

Disposal must comply with national recommendations and must take into account the internal requirements of the medical facility.

### **15. Symbols used**



Symbol for "Item Number"



Symbol for "Batch code"



Symbol for "Manufacturer"



Symbol for "Date of manufacture"



Symbol for "Observe instructions for use"



Symbol for "Sterilized with ethylene oxide"



Symbol for "non-reuse"



Symbol for "Do not re-sterilize"



Symbol for "Do not use if packaging is damaged"



Symbol for "Expiry Date"



Symbol for "keep dry"



Symbol for "protect from sunlight"



Symbol for "Attention"



Symbol for "Medical Device"



Symbol for "simple sterile barrier system"



Symbol for "simple sterile barrier system with outer protective packaging"